

## Summary results

### BLOT-LINE ANA

					Summary results BLOT-LINE ANA								
Lot: A0199324 Performed by: VENNELA Evaluation time: 4/8/2024 10:55:55 AM					Borderline area from 8.00 to 10.00								
Scan image(1 - 19):													
	Sample no. 1	Sample no. 2	Sample no. 3	Sample no. 4	Sample no. 6 (#+)	Sample no. 7	Sample no. 9	Sample no. 17 (+)	Sample no. 18	Sample no. 19			
Results(1 - 19):													
	Sample no. 1	Sample no. 2	Sample no. 3	Sample no. 4	Sample no. 6 (#+)	Sample no. 7	Sample no. 9	Sample no. 17 (+)	Sample no. 18	Sample no. 19			
Description											Ms. SHIVANI PASI		
<b>IgG Control</b>	✓	✓	✓	✓	X	✓	✓	✓	✓	✓	X		
<b>Jo-1</b>	0.51[N]	0.00[N]	0.00[N]	0.99[N]	<b>25.92[P]</b>	0.97[N]	0.47[N]	1.96[N]	0.00[N]	0.00[N]			
<b>PL-7</b>	1.01[N]	0.00[N]	1.01[N]	0.00[N]	3.04[N]	0.97[N]	0.47[N]	0.99[N]	0.00[N]	0.50[N]			
<b>PL-12</b>	1.52[N]	0.00[N]	1.51[N]	0.50[N]	3.50[N]	1.45[N]	0.00[N]	2.46[N]	1.01[N]	1.01[N]			
<b>PmScl</b>	2.02[N]	0.00[N]	0.51[N]	1.00[N]	2.19[N]	0.49[N]	0.00[N]	<b>11.05[P]</b>	0.51[N]	1.51[N]			
<b>CENP-A</b>	2.02[N]	0.00[N]	3.50[N]	0.00[N]	2.63[N]	0.49[N]	0.00[N]	3.44[N]	0.51[N]	2.00[N]			
<b>CENP-B</b>	0.51[N]	0.00[N]	0.00[N]	2.48[N]	2.61[N]	0.00[N]	0.47[N]	0.00[N]	0.00[N]	0.50[N]			
<b>Scl-70</b>	0.00[N]	1.01[N]	0.00[N]	0.00[N]	1.31[N]	0.00[N]	0.47[N]	2.48[N]	1.01[N]	1.50[N]			
<b>Ro52</b>	<b>16.82[P]</b>	<b>44.98[P]</b>	<b>30.69[P]</b>	<b>40.90[P]</b>	0.44[N]	1.94[N]	3.26[N]	1.49[N]	1.51[N]	1.51[N]			
<b>Ro60</b>	<b>12.13[P]</b>	3.99[N]	<b>69.19[P]</b>	<b>40.47[P]</b>	0.44[N]	0.98[N]	3.26[N]	2.97[N]	0.00[N]	1.01[N]			
<b>La</b>	<b>33.32[P]</b>	<b>56.29[P]</b>	<b>72.10[P]</b>	<b>46.02[P]</b>	3.94[N]	0.00[N]	0.47[N]	<b>11.05[P]</b>	0.00[N]	0.00[N]			
<b>RNP-A</b>	0.00[N]	5.44[N]	5.92[N]	1.99[N]	0.88[N]	0.00[N]	0.94[N]	<b>18.41[P]</b>	1.01[N]	0.50[N]			
<b>RNP-C</b>	0.00[N]	2.50[N]	7.36[N]	0.00[N]	2.20[N]	0.00[N]	0.47[N]	<b>19.67[P]</b>	0.00[N]	0.00[N]			
<b>RNP-68</b>	0.00[N]	1.98[N]	0.50[N]	1.00[N]	2.62[N]	1.47[N]	0.47[N]	<b>13.34[P]</b>	1.01[N]	1.00[N]			
<b>SmB</b>	0.00[N]	2.49[N]	<b>11.62[P]</b>	0.00[N]	1.31[N]	1.47[N]	0.00[N]	9.15[B]	1.01[N]	0.50[N]			
<b>SmD</b>	3.02[N]	6.86[N]	<b>15.80[P]</b>	1.00[N]	1.30[N]	0.49[N]	0.47[N]	<b>32.55[P]</b>	0.00[N]	7.81[N]			
<b>P0</b>	2.52[N]	0.00[N]	1.01[N]	1.01[N]	0.00[N]	0.00[N]	0.00[N]	2.95[N]	1.01[N]	0.00[N]			
<b>PCNA</b>	0.00[N]	0.00[N]	0.00[N]	1.01[N]	0.00[N]	0.00[N]	0.00[N]	0.00[N]	0.00[N]	0.00[N]			
<b>dsDNA</b>	1.02[N]	0.00[N]	6.88[N]	8.76[B]	1.31[N]	0.98[N]	1.90[N]	<b>76.65[P]</b>	1.00[N]	2.49[N]			
<b>Histone</b>	0.00[N]	3.97[N]	2.99[N]	1.00[N]	9.55[B]	0.98[N]	0.95[N]	<b>70.02[P]</b>	0.50[N]	0.50[N]			
<b>Nucleosom</b>	1.02[N]	1.51[N]	2.00[N]	1.51[N]	0.44[N]	5.78[N]	3.79[N]	<b>53.72[P]</b>	2.49[N]	1.00[N]			
<b>DFS70</b>	0.00[N]	1.01[N]	0.00[N]	0.00[N]	<b>16.68[P]</b>	0.00[N]	0.48[N]	<b>12.41[P]</b>	0.50[N]	0.00[N]			
<b>M2</b>	0.00[N]	0.51[N]	0.00[N]	1.01[N]	0.87[N]	0.00[N]	0.00[N]	1.48[N]	0.50[N]	0.50[N]			
<b>ANA</b>	<b>Positive</b>	<b>Positive</b>	<b>Positive</b>	<b>Positive</b>	<b>Positive</b>	Negative	Negative	<b>Positive</b>	Negative	Negative			

Strips assigned (+) - In case of negative result of ANA and positive result of DFS70: low probability of SARD. In case of positive result of M2: potential risk of primary biliary cirrhosis progression.

Strips assigned (#) - Control line located incorrectly, SW evaluation result is not valid

The evaluation has been accomplished without calibration of the input image

X - Conjugate control line was not found. Please check the possibility of substitution

## Summary results

### BLOT-LINE ANA

<b>Lot:</b> A0199324 <b>Performed by:</b> VENNELA <b>Evaluation time:</b> 4/8/2024 10:55:55 AM		<b>Borderline area from 8.00 to 10.00</b>									
		<b>Scan image(20 - 20):</b>									
	<b>Sample no. 20</b>										
<b>Results(20 - 20):</b>											
	<b>Sample no. 20</b>										
<b>Description</b>	Mrs. GEETASHRI KANHERE										
<b>IgG Control</b>	✓										
<b>Jo-1</b>	0.00[N]										
<b>PL-7</b>	0.51[N]										
<b>PL-12</b>	0.51[N]										
<b>PmScl</b>	0.00[N]										
<b>CENP-A</b>	0.51[N]										
<b>CENP-B</b>	0.00[N]										
<b>Scl-70</b>	0.51[N]										
<b>Ro52</b>	0.00[N]										
<b>Ro60</b>	1.02[N]										
<b>La</b>	0.00[N]										
<b>RNP-A</b>	0.00[N]										
<b>RNP-C</b>	0.00[N]										
<b>RNP-68</b>	0.00[N]										
<b>SmB</b>	0.51[N]										
<b>SmD</b>	0.00[N]										
<b>P0</b>	0.00[N]										
<b>PCNA</b>	0.00[N]										
<b>dsDNA</b>	3.53[N]										
<b>Histone</b>	1.02[N]										
<b>Nucleosom</b>	2.02[N]										
<b>DFS70</b>	0.00[N]										
<b>M2</b>	1.52[N]										
<b>ANA</b>	Negative										
Strips assigned (+) - In case of negative result of ANA and positive result of DFS70: low probability of SARD. In case of positive result of M2: potential risk of primary biliary cirrhosis progression.											
Strips assigned (#) - Control line located incorrectly, SW evaluation result is not valid											
The evaluation has been accomplished without calibration of the input image											
X - Conjugate control line was not found. Please check the possibility of substitution											