


A2181338	
Mr. ANKUR	02
Director:	Panel: HLA-B27 Acquired: 05-07-2025 15:02:00 Analyzed: 05-07-2025 15:47:00 Bead Lot ID: 1231345/139 Reagent Lot ID: 1299250:147 Operator: J. RAVI Results: 05072025.csv
Column 1#: 15 YRS/Male	

BD FACSCanto II V33896202252

BD FACSCanto v.3.0.4894.41215

HLA-B27

Total Events: 27891



Reagent Lot ID: 1299249:146

Gated Events	2783
Preset HLA-B27 Marker	175
Sample HLA-B27 Median	177
Conclusion	HLA-B27 Positive Sample

QC Messages

No quality control messages generated.

Comments