


A2209290	
Mr. ATUL KUMAR	02
Director:	Panel: HLA-B27 Acquired: 09-07-2025 15:51:00 Analyzed: 09-07-2025 16:26:00 Bead Lot ID: 1231345/139 Reagent Lot ID: 1299250:147 Operator: J. RAVI Results: 09072025.csv
Column 1#: 26 YRS/Male	

BD FACSCanto II V33896202252

BD FACSCanto v.3.0.4894.41215

HLA-B27

Total Events: 28619



Reagent Lot ID: 1299249:146

Gated Events	2853
Preset HLA-B27 Marker	175
Sample HLA-B27 Median	177
Conclusion	HLA-B27 Positive Sample

QC Messages
 No quality control messages generated.

Comments