


A1889774	
Ms. KALAWATI BAI	01
Director:	Panel: HLA-B27 Acquired: 18-07-2025 19:04:00 Analyzed: 18-07-2025 19:04:00 Bead Lot ID: 1231345/139 Reagent Lot ID: 1299250:147 Operator: J. RAVI Results: 18072025.csv
Column 1#: 38 YRS/Female	

BD FACSCanto II V33896202252

BD FACSCanto v.3.0.4894.41215

HLA-B27

Total Events: 30422



Reagent Lot ID: 1299249:146

Gated Events	3036
Preset HLA-B27 Marker	199
Sample HLA-B27 Median	204
Conclusion	HLA-B27 Positive Sample

QC Messages

No quality control messages generated.

Comments